ZERTIFIKAT ◆ CERTIFICATE ◆ 認証書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Product Service

CERTIFICATE

No. Z1 10 09 62765 003

Holder of Certificate: **AQUARIA S.R.L.**
Via della Levata, 14
20084 Lacchiarella MI
ITALY

Production Facility(ies): 62766

Certification Mark:

TÜV SÜD
Production monitored
Safety tested

Product: **Laboratory equipment (Microbiological safety cabinet)**

Model(s): **BIO ACTIVA 180, BIO ACTIVA VE 180, FLOW ACTIVA VF 180_PLUS, FLOW ACTIVA VE 180_PLUS, BIO ACTIVA ONE 180, CYTO ACTIVA 180, CYTO ACTIVA VE 180, BIO ACTIVA VE 150, FLOW ACTIVA VF 150_PLUS, FLOW ACTIVA VE 150_PLUS, BIO ACTIVA ONE 150, CYTO ACTIVA VE 150**

Parameters:

Rated voltage:	230 V
Rated input power:	See attachment
Max power on socket outlet:	800 W
Class of MSC:	II
Safe working area:	See attachment

Tested according to: EN 61010-1:2001
EN 12469:2000

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: SIC10035.01

Date, 2010-09-30 (Emanuele Ferrari)

369341

ZERTIFIKAT ◆ CERTIFICATE ◆ 認証書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Product Service

Attachment to

CERTIFICATE

No. Z1 10 09 62765 003

Model	Safe working area	Rated Input Power
BIO ACTIVA 180	1795 x 600 mm	1000 W
BIO ACTIVA VE 180	1825 x 690 mm	1000 W
FLOW ACTIVA VF 180_PLUS	1880 x 570 mm	700 W
FLOW ACTIVA VE 180_PLUS	1898 x 570 mm	700 W
BIO ACTIVA ONE 180	1880 x 690 mm	1000 W
CYTO ACTIVA VE 180	1790 x 600 mm	1000 W
CYTO ACTIVA 180	1825 x 690 mm	1000 W

Model	Safe working area	Rated Input Power
BIO ACTIVA VE 150	1400 x 690 mm	670 W
FLOW ACTIVA VF 150_PLUS	1460 x 570 mm	670 W
FLOW ACTIVA VE 150_PLUS	1460 x 570 mm	670 W
BIO ACTIVA ONE 150	1455 x 690 mm	670 W
CYTO ACTIVA VE 150	1400 x 690 mm	1000 W

Date, 2010-09-30

Product Service

TÜV SÜD Product Service GmbH · Zertifizierstelle · Ridlerstrasse 65 · 80339 München · Germany

TÜV®