ZERTIFIKAT ◆ CERTIFICATE ◆ 認証書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Product Service

CERTIFICATE

No. Z1 10 09 62765 004

Holder of Certificate: **AQUARIA S.R.L.**
Via della Levata, 14
20084 Lacchiarella MI
ITALY

Production Facility(ies): 62766

Certification Mark:

TÜV SÜD
Production monitored
Safety tested

Product: Laboratory Equipment
(Microbiological safety cabinet)

Model(s):
BIO ACTIVA VE 120,
BIO ACTIVA ONE 120,
BIO ACTIVA 120,
FLOW ACTIVA VF 120_PLUS,
FLOW ACTIVA VE 120_PLUS,
CYTO ACTIVA VE 120,
CYTO ACTIVA 120.

Parameters:

Rated voltage:	230 V
Rated input power:	660 W
Man power on socket outlet:	800 W
Class of MSC:	II
Safe working area:	See attachment

Tested according to:
EN 61010-1:2001
EN 12469:2000

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: SIC06046.03

Date, 2010-09-08

(Fabrizio Gasparella)

Page 1 of 2

TÜV Product Service
369284

A1 / 03.08

TÜV®

ZERTIFIKAT ◆ CERTIFICATE ◆ 認証書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Product Service

Attachment to

CERTIFICATE

No. Z1 10 09 62765 004

Model	Safe working area
BIO ACTIVA VE 120	1220 x 690 mm
BIO ACTIVA ONE 120	1275 x 690 mm
BIO ACTIVA 120	1185 x 600 mm
FLOW ACTIVA VF 120_PLUS	1275 x 570 mm
FLOW ACTIVA VE 120_PLUS	1275 x 570 mm
CYTO ACTIVA VE 120	1220 x 690 mm
CYTO ACTIVA 120	1185 x 600 mm

Product Service

Date, 2010-09-10

TÜV SÜD Product Service GmbH · Zertifizierstelle · Ridlerstrasse 65 · 80339 München · Germany

TÜV®